



NATIONAL PHYSICAL LABORATORY

Teddington Middlesex UK TW11 0LW Telephone +44 20 8977 3222

Certificate of Calibration



4002

NPL PRIMARY REFERENCE MATERIAL

Cylinder Number: 2795

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CUSTOMER: Deutscher Wetterdienst
ADDRESS: Meteorologisches Observatorium Hohenpeißenberg, Albin-Schwaiger-Weg 10, 82383 Hohenpeißenberg, Germany
CALIBRATION DATE: 30 August 2019

AMOUNT FRACTION:

Component	Amount fraction / ($\mu\text{mol/mol}$)
Nitrogen monoxide	10.014 ± 0.040
Nitrogen	Balance

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a coverage probability of approximately 95 %. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

METHODS: Preparation: gravimetry; Analysis: chemiluminescence
TRACEABILITY: The values on this certificate are traceable to NPL Primary Standards
EXPIRY: Certificate valid for 2 years from the date of issue
PRESSURE: Fill pressure: 110 bar; Minimum utilisation pressure: 5 bar
STORAGE: No special precautions are required
HANDLING: Refer to ISO 16664
OUTLET: BS341 No. 14 valve
INTENDED USE: Calibration standard

Reference: 2018110394

Date of issue: 30 August 2019

Signed:  (Authorised Signatory)

Name: Dr R J C Brown (on behalf of NPLML)

Checked by: 

Page 1 of 1



This certificate is consistent with the capabilities that are included in Appendix C of the MRA drawn up by the CIPM. Under the MRA, all participating institutes recognise the validity of each other's calibration and measurement certificates for the quantities, ranges and measurement uncertainties specified in Appendix C (for details see <http://www.bipm.org>).